

Applicant: CIRCLE SQUARE HOLDINGS LTD
7 BELL YARD, HOLBORN, LONDON,
WC2A 2JR, UNITED KINGDOM

Date: Sep 30, 2020

Attn: ANDREW PRINGLE

Sample Description:

Four hundred (400) pieces of submitted samples said to be Nitrile Medical gloves in Blue.

| | | |
|-----------------------------------|---|--|
| Standard | : | EN 455-1:2000 |
| | | EN 455-2:2015 |
| | | EN 455-3:2015 |
| P.O. No. | : | PO-0015 |
| Colors | : | Blue |
| Size Range | : | S, M, L |
| Style Name | : | Sri Trang-Balance Nitrile Gloves |
| Buyer's Name | : | MedicShield (Trading name of Circle Square Holdings Ltd) |
| Vendor | : | MedicShield (Trading name of Circle Square Holdings Ltd) |
| Supplier | : | Henan Tongzhikang Medical Devices Co Ltd |
| Manufacturer | : | Sri Trang-Balance |
| Country Of Origin | : | Thailand |
| Goods Exported To | : | United Kingdom |
| Date Received/Date Test Started : | | Sep 21, 2020/ Sep 21, 2020 |
| Date Final Information Confirmed/ | | --/-- |
| Date Payment Received: | | |

Test Result Please Refer To Attached Page(S).

Should you have any query on this report, you may contact at gzfootwear@intertek.com

Authorized By:
For Intertek Testing Services Shenzhen Ltd.
Guangzhou Branch



Guiliang Dong
Senior Lab Manager



MI/fionawgyu

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F. & Room 01, E101/E201/E301/E401/E501/E601/E701/E801, No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, Guangdong, China
3/F., Hengyun Building, 235 Kaifa Ave., Guangzhou Economic & Technological Development District, Guangzhou, China
广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房101、E201、E301、E401、E501、E601、E701、E801
中国广州经济技术开发区开发大道235号恒运大厦3楼
Tel: +86 208213 9001 Fax: +86 20 82089909 Postcode: 510663
Tel: +86 20 83966868 Fax: +86 20 82228169 Postcode: 510730



1 Watertightness Test For Detection Of Holes (EN 455-1:2000, 5)

| Results | | | Requirement | | | Pass/Fail |
|----------------------|---------|-------|----------------------|-----|-----------|-----------|
| Physical Performance | Failure | AQL | Physical Performance | AQL | n [Ac Re] | - |
| Water Leakage | 2 | < 1.5 | No Leakage | 1.5 | 200 [7,8] | Pass |

Remark: n Means Sample Size, Ac Means Acceptance Number, Re Means Rejection Number.

2 Dimensions Of Examination Gloves (BS EN 455-2:2015 / EN 455-2:2015, 4)

| Sample | Size | Measurement | | Requirement | Pass/Fail |
|--------|--------|--------------------|-------------------|-------------|-----------|
| | | Median Length (mm) | Median Width (mm) | | |
| - | Small | 245 | 85 | * | Pass |
| | Medium | 245 | 97 | * | Pass |
| | Large | 242 | 107 | * | Pass |

Remark: *= Dimensions Of Examination / Procedure Gloves

| Size | Median Length (mm) | Median Width (mm) |
|---|--------------------|-------------------|
| Extra Small | ≥ 240 | ≤ 80 |
| Small | | 80 ± 10 |
| Medium | | 95 ± 10 |
| Large | | 110 ± 10 |
| Extra Large | | ≥ 110 |
| Note: The Width Requirements Are For Gloves Made From Natural Rubber Latex And All Other Elastomeric Material. These Dimensions May Not Appropriate For Gloves Made From Other Materials. | | |



3 Force At Break (EN 455-2:2015, 5.2 & 5.3 & ISO 188:2007)

| Total Number Tested | Result | Requirement | Pass/Fail |
|---------------------|--|-------------|-----------|
| 13 | Before Challenge Testing Median Values Of Force At Break | Min. 6.0 N | Pass |
| | 6.5N | | |
| 13 | After Challenge Testing (70 °C, 7 Days) Median Values Of Force At Break | Min. 6.0 N | Pass |
| | 6.3N | | |

Remark:

Requirements Of Median Values Of Force At Break

| | Force At Break (N) | | |
|---|--------------------|--------------------------------|------------|
| | Surgical Gloves | Examination / Procedure Gloves | |
| | a) | b) | c) |
| Throughout Shelf Life Tested According To 5.2 And Within 12 Months Of Manufacture Tested According To 5.3 | ≥ 9.0 | ≥ 6.0 | ≥ 3.6 |
| a) Requirements For All Surgical Gloves. | | | |
| b) Requirements For All Examination Gloves, Except Gloves Made From Thermoplastic Materials (e.g. Polyvinylchloride, Polyethylene). | | | |
| c) Requirements For Gloves Made From Thermoplastic Materials (e.g. Polyvinylchloride, Polyethylene). | | | |

4 Powder-free Gloves (EN 455-3:2015, 5.2 & EN ISO 21171:2006, Method B/ Method C)

| Result | Requirement | Pass/Fail |
|--------------|---------------------|-----------|
| 0.1 mg/Glove | ≤ 2.0 mg/Glove | Pass |

MI / fionawgyu

Page 3 Of 6

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F. & Room 01, E101/E201/E301/E401/E501/E601/E701/E801
No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, Guangdong, China
广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房101、
E201、E301、E401、E501、E601、E701、E801

Tel: +86 208213 9001 Fax: +86 20 82089909 Postcode: 510663

3/F., Hengyun Building, 235 Kaifa Ave., Guangzhou
Economic & Technological Development District, Guangzhou,
China

中国广州经济技术开发区开发大道235号恒运大厦3楼

Tel: +86 20 83966868 Fax: +86 20 82228169 Postcode: 510730



5 Chemicals (EN 455-3:2015, 4.2)

| Requirements | Pass | Fail | N/A |
|---|------|------|-----|
| Gloves Shall Not Be Dressed With Talcum Powder (Magnesium Silicate). | X | | |
| The Manufacturer Shall Disclose, Upon Request, A List Of Chemical Ingredients Either Added During Manufacturing Or Already Known To Be Present In The Product Such As Accelerators, Antioxidants And Biocides That Are Known To Cause Adverse Health Effects Based On Current Data. | X | | |
| Upon Request The Manufacturer Shall Provide Evidence Of The Steps Taken To Reduce The Risk To The End-User Of Exposure To Chemicals Used In The Manufacturing Process Which, Based On Current Data, Are Known To Cause Adverse Health Effects. | X | | |
| Manufacturers May Only Declare The Absence Of A Substance If The Substance Is Not Used In Any Part Of The Manufacturing Process. No Compounds Shall Be Used In The Manufacture Of The Product That Is Known To Form A Substance That Is Subject Of Such A Declaration. | X | | |

Compliance: The Submitted Sample **MEETS** The Requirements Of EN 455-3:2015 Clause 4.2 For Chemicals.



6 Labelling (EN 455-3:2015, 4.6)

| Requirements | | Pass | Fail | N/A |
|---|---|------|------|-----|
| Labelling Specified In EN 1041:2008+A1:2013 And The Relevant Symbols Given In EN ISO 15223-1:2012: | | | | |
| 1) | Description Of Sample | X | | |
| 2) | Name And Address Of Manufacturer | X | | |
| 3) | Date Of Manufacture | X | | |
| 4) | Use By Date | X | | |
| 5) | Batch Code/ Lot Number/ Serial Number | X | | |
| 6) | Relevant Information Of Sterility | | | X |
| 7) | Relevant Information Of Single Use | X | | |
| 8) | Storage conditions | X | | |
| In Addition To The Labelling Specified In EN 1041:2008+A1:2013 And The Relevant Symbols Given In EN ISO 15223-1:2012, The Following Requirements Apply: | | | | |
| a) | Medical Gloves Containing Natural Rubber Latex Shall Be Labelled On The Packaging Of At Least The Smallest Packaging Unit With The EN ISO 15223-1:2012 Symbol For Latex (Reference Number 5.4.5). The Labelling Shall Include The Following Or Equivalent Warning Statement Together With The Symbol (Product) Contains Natural Rubber Latex Which May Cause Allergic Reactions, Including Anaphylactic Responses; | | | X |
| b) | The Labelling Shall Include A Prominent Indication Of Whether The Glove Is Powdered Or Powder-Free; | X | | |
| c) | Sterile Powdered Gloves Shall Be Labelled With The Following Or Equivalent: 'CAUTION: Surface Powder Shall Be Removed Aseptically Prior To Undertaking Operative Procedures In Order To Minimize The Risk Of Adverse Tissue Reactions'; (NOTE 1 This Caution Statement Can Be Given On The Inner Wrapping.) | | | X |
| d) | For Any Medical Glove Containing Natural Rubber Latex The Product Labelling Shall Not Include: - Any Term Suggesting Relative Safety, Such As Low Allergenicity, Hypoallergenicity Or Low Protein; - Any Unjustified Indication Of The Presence Of Allergens; | | | X |
| e) | If The Manufacturer Labels The Gloves With The Protein Content, The Process Limit, Measured As Specified In 5.3 Shall Be Given. (NOTE 2 This Does Not Allow A Protein Labelling Claim Below 50 µG/G. Lower Claims Are Not Considered To Be Reliable Given The Expected Process Variation In Manufacture And Inter-Laboratory Testing.) | | | X |

Compliance: The Submitted Sample **MEETS** The Requirements Of EN 455-3:2015 Clause 4.6 For Labelling.





End of Report

This report is made solely on the basis of your instructions and/or information and materials supplied by you. It is not intended to be a recommendation for any particular course of action. Intertek does not accept a duty of care or any other responsibility to any person other than the Client in respect of this report and only accepts liability to the Client insofar as is expressly contained in the terms and conditions governing Intertek's provision of services to you. Intertek makes no warranties or representations either express or implied with respect to this report save as provided for in those terms and conditions. We have aimed to conduct the Review on a diligent and careful basis and we do not accept any liability to you for any loss arising out of or in connection with this report, in contract, tort, by statute or otherwise, except in the event of our gross negligence or wilful misconduct. No copy of the test report(except for full text copy) shall be made without the written approval by Intertek.

MI / fionawgyu

Page 6 Of 6

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

深圳天祥质量技术服务有限公司广州分公司

Room 02, 1-8/F. & Room 01, E101/E201/E301/E401/E501/E601/E701/E801

No.7-2, Caipin Road, Guangzhou Science City, GETDD, Guangzhou, Guangdong, China

广州经济技术开发区科学城彩频路7号之二第1-8层02房、01房101、

E201、E301、E401、E501、E601、E701、E801

Tel: +86 208213 9001 Fax: +86 20 82089909 Postcode: 510663

3/F., Hengyun Building, 235 Kaifa Ave., Guangzhou

Economic & Technological Development District, Guangzhou,

China

中国广州经济技术开发区开发大道235号恒运大厦3楼

Tel: +86 20 83966868 Fax: +86 20 82228169 Postcode: 510730

